

Instrument
Overload Status
OVLS? [i]
OVLC? [i]
OVLE? [i]
OVLE {j}
*CLS
Bits
Bytes
Wgt
Pos.
Label
OVLC
OVLS
OVLE
128 7 rfu R R R/W
64 6 rfu R R R/W
32 5 rfu R R R/W
16 4 CLN R R R/W
8 3 CLP R R R/W
4 2 FBK R R R/W
2 1 REF R R R/W
1 0 PCB R R R/W
Access
OR
Instrument Status
INSS? [i]
INSC? [i]
INSE? [i]
INSE {j}
*CLS
INSC
INSS
INSE
128 7 rfu R R R/W
64 6 CRG R R R/W
32 5 ERG R R R/W
16 4 FRG R R R/W
8 3 IPW R R R/W
4 2 LKD R R R/W
2 1 IKS R R R/W
1 0 PUV R R R/W
Last Instrument Error
LINS? *CLS
Byte
Access
Value
LINS R 0-255
Last User Request
LURQ? *CLS
LURQ R 0-255
Master Summary Status
MSTS? [i]
MSTE? [i]
MSTE {j}
*CLS
MSTC
MSTS
MSTE
128 7 OVL N/A R R/W
64 6 INS N/A R R/W
32 5 RFU N/A R R/W
16 4 RFU N/A R R/W
8 3 RFU N/A R R/W
4 2 EVT N/A R R/W
2 1 COM N/A R R/W
1 0 MSS N/A R
assert
de-assert
MSTS?
/Status
Remote Control
Last Execution Error
LEXE? *CLS
LEXE R 0-255
Last Command Error
LCMD? *CLS
LCMD R 0-255
Communications Status
Not used
COMS? [i]
COME? [i]
COME {j}
*CLS
COMC
COMS
COME
128 7 RFU N/A R R/W
64 6 RFU N/A R R/W
32 5 RFU N/A R R/W
16 4 RFU N/A R R/W
8 3 RFU N/A R R/W
4 2 RFU N/A R R/W
2 1 COL N/A R R/W
1 0 PRY N/A R R/W
Event Summary Status
EVTS? [i]
EVTE? [i]
EVTE {j}
*CLS
EVTC
EVTS
EVTE
128 7 INS N/A R R/W
64 6 URQ N/A R R/W
32 5 TXQ N/A R R/W
16 4 RXQ N/A R R/W
8 3 EXE N/A R R/W
4 2 CMD N/A R R/W
2 1 OPC N/A R R/W
1 0 PON N/A R R/W
*OPC
Legend
R/W Read/Write value
R Read-Only value
RFU Reserved for future use
N/A Not Applicable or Not Available
Signals & Systems for Physics
www.sisyph.com
206 Rue Alfred Kastler
17190 Saint-Georges-d'Oléron
FRANCE
Title/Name
SK633 Sancy - Low-Noise Phase-Lock Controller
Release
SK305-R26A
Description
Status Register Model
Drawing No.
SK633-SG02-P26A
Sheet Size
A4
Sheet No.
1 of 1
APP'D By
DRN By
GGB
26-11-06
Name
Date
YY-MM-DD
THIS DOCUMENT IS THE PROPERTY OF SISYPH AND SHALL NOT TO BE REPRODUCED, DISTRIBUTED OR USED WITHOUT PRIOR WRITTEN PERMISSION FROM SISYPH.
PRELIMINARY INFORMATION